

GMT221 and GMT222 CO₂ Transmitters

| Vaisala CARBOCAP® CO ₂ Transmitter | | GM | T | 22 | | | | | | | | | | | | | | PRICE |
|---|--|---------------------------------|---|----|---|---|--|--|--|--|--|--|--|--|--|-------------|--|-------|
| Measurement range (GMT221) | 0...2 %CO₂ | | | | 1 | H | | | | | | | | | | | | |
| | 0...3 %CO ₂ | | | | 1 | J | | | | | | | | | | | | |
| | 0...5 %CO ₂ | | | | 1 | K | | | | | | | | | | | | |
| | 0...10 %CO ₂ | | | | 1 | L | | | | | | | | | | | | |
| | 0...20 %CO ₂ | | | | 1 | M | | | | | | | | | | | | |
| (GMT222) | 0...2000 ppmCO₂ adjusted at 1000 ppm (* | | | | 2 | B | | | | | | | | | | | | |
| | 0...2000 ppmCO ₂ | | | | 2 | C | | | | | | | | | | | | |
| | 0...3000 ppmCO ₂ | | | | 2 | D | | | | | | | | | | | | |
| | 0...5000 ppmCO ₂ | | | | 2 | E | | | | | | | | | | | | |
| | 0...7000 ppmCO ₂ | | | | 2 | F | | | | | | | | | | | | |
| | 0...10000 ppmCO ₂ | | | | 2 | G | | | | | | | | | | | | |
| | Temperature compensation | standard setting (25 °C) | | | | 0 | | | | | | | | | | | | |
| | special setting (37 °C) | | | | 1 | | | | | | | | | | | | | |
| Probe installation accessories | none | | | | | | | | | | | | | | | | | |
| | <i>mounting clips, 2 pcs</i> | | | | | | | | | | | | | | | | | |
| | <i>mounting flange</i> | | | | | | | | | | | | | | | | | |
| Field check adapter | none | | | | 0 | | | | | | | | | | | | | |
| | <i>adapter</i> | | | | 1 | | | | | | | | | | | | | |
| Display on cover | LED lights only | | | | | | | | | | | | | | | | | |
| | <i>LED lights + backlit digital LCD display</i> | | | | | | | | | | | | | | | | | |
| Wall/remote use | none | | | | 0 | | | | | | | | | | | | | |
| | probe in transmitter (wall model) | | | | | | | | | | | | | | | | | |
| | <i>probe with cable (remote installation of probe)</i> | | | | | | | | | | | | | | | | | |
| | <i>probe with 10 m cable (remote installation of probe)</i> | | | | | | | | | | | | | | | | | |
| Alarm relay 1 | setpoint not activated | | | | | | | | | | | | | | | | | |
| | custom setpoint for relay 1 _____ ppm / % CO₂ | | | | | | | | | | | | | | | | | |
| | Example: custom setpoint for relay 1 <u>1000</u> ppm / % CO ₂ | | | | | | | | | | | | | | | | | |
| | -> Relay will be activated above 1000 ppm CO ₂ | | | | | | | | | | | | | | | | | |
| Alarm relay 2 | setpoint not activated | | | | | | | | | | | | | | | | | |
| | custom setpoint for relay 2 _____ ppm / % CO₂ | | | | | | | | | | | | | | | | | |
| | Example: custom setpoint for relay 2 <u>4</u> ppm / % CO ₂ | | | | | | | | | | | | | | | | | |
| | -> Relay will be activated above 4% CO ₂ | | | | | | | | | | | | | | | | | |
| Accessories | none | | | | 0 | | | | | | | | | | | | | |
| | <i>serial COM adapter</i> | | | | 1 | | | | | | | | | | | | | |
| Manual | No manual | | | | | | | | | | | | | | | | | |
| | English | | | | | | | | | | | | | | | | | |
| | Japanese | | | | | | | | | | | | | | | | | |
| | | | | | | | | | | | | | | | | TOTAL | | |
| | | | | | | | | | | | | | | | | QTY | | |
| | | | | | | | | | | | | | | | | TOTAL VALUE | | |

Selections in bold are included in the prices of the basic versions.
Selections in italic are available at an extra price.

(* special accuracy specification)

Examples of order codes with typical settings:

Transmitter

GM|T|22|1|L|0|N|0|B|0|N|1|A|0|B

End customer: _____

Accessories:

- Spare membrane filter plus cap for GMP221
- Spare membrane filter plus cap for GMP222
- Protective sleeve for the GMP221
- Protective sleeve for the GMP222
- In-soil adapter kit

Order code:

- 25378GMSP
- 25879GMSP
- GM45168SP
- GM45237SP
- 211921GM